

AUG 03 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

| | |
|--------------------------------------------------------------------|-------------------------------------------------------------------------------------------|
| Supervisor/Operator Performing the Verification Procedure: | |
| Name <u>JAMES MICHEL</u> | ID# <u>6502</u> Date <u>8-1-10</u> ✓ |
| DataMaster cdm S/N <u>130129</u> ✓ | |
| A | Agency <u>USB POLICE DEPT</u> Phone # <u>852-0311</u> |
| Instrument Location <u>BARROW - SQUAD ROOM</u> | |
| B | Alco S/N <u>X301568</u> ✓ Target Value <u>.0841</u> ✓ High Pressure <u>850</u> <u>psi</u> |
| Alco Test Values <u>.080</u> ✓ <u>.0841</u> ✓ 1st Alco 2nd Alco | |
| Signature <u>[Signature]</u> SGT. MICHEL SFS | |
| (OVER) | |

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 11th day of AUGUST, 2010.

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130129

(CONTINUED FROM FRONT PAGE)

CB
8/9/10

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

AUGUST 01, 2010

OPERATOR'S NAME:

MICHELSON, JAMES/D

OPERATOR'S NUMBER: 6502

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: BRW1

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .004

ALCO S/N: X301568

--- BREATH ANALYSIS ---

| | | |
|----------------------------|----------|-------|
| .004 ADJUSTED FOR 29.93 in | | |
| ALCO TARGET | .004 | 17:13 |
| BLANK TEST | .000 | 17:14 |
| INTERNAL STANDARD | VERIFIED | 17:14 |
| ALCO TV 29.93 in | .000 | 17:14 |
| BLANK TEST | .000 | 17:15 |
| SUBJECT SAMPLE | .000 | 17:15 |
| BLANK TEST | .000 | 17:16 |
| ALCO TV 29.93 in | .004 | 17:16 |
| BLANK TEST | .000 | 17:17 |

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

AUGUST 01, 2010

TIME 17:11

--- DIAGNOSTIC CHECK ---

| | |
|------------------|----------|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 01/08/09 |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| BREATH TUBE: | 42c |
| BAROMETER: | 29.93 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~